Month	Jan	Feb	Mar	Apr	May	Jun	Jul	Aug	Sep	Oct	Nov	Dec
Reading in iced water												
Reading in boiling water												
Checked by												

- The readings in **iced water should be -1°C to +1°C**, if outside this range the unit should be replaced or returned to the manufacturer to be recalibrated.
- The reading in **boiling water should be between 99°C and 101°C**, if outside this range the unit should be replaced or returned to the manufacturer to be recalibrated.

Date and Details of Yearly Calibration		**Date of next Yearly Calibration**	

N.B. The electronic display unit should be checked at least once per year. Manufacturers may offer a calibration service.

Month	Jan	Feb	Mar	Apr	May	Jun	Jul	Aug	Sep	Oct	Nov	Dec
Reading in iced water												
Reading in boiling water												
Checked by												

- The readings in **iced water should be -1°C to +1°C**, if outside this range the unit should be replaced or returned to the manufacturer to be recalibrated.
- The reading in **boiling water should be between 99°C and 101°C**, if outside this range the unit should be replaced or returned to the manufacturer to be recalibrated.

Date and Details of Yearly Calibration		**Date of next Yearly Calibration**	

N.B. The electronic display unit should be checked at least once per year. Manufacturers may offer a calibration service.